



## 1. GENERAL

This unit is a transistorized dip meter for rapidly checking circuits and components in receivers, transmitters, antennas, etc., in the 1.5 to 250MHz range. It is designed for many applications in Amateur Radio stations and service shops.

## 2. SPECIFICATIONS

Frequency Range	: 1.5 to 250MHz with six plugin coils:
	Band                      Range
	A                      1.5 – 4MHz
	B                      3.3 – 8MHz
	C                      6.8 – 18MHz
	D                      18 – 47MHz
	E                      45 – 110MHz
	F                      100 – 250MHz
Modulation	: Approx. 2kHz; sine wave
Crystal Oscillator	: 1-15MHz, crystal in FT-243 holder
Power Supply	: 9 volt battery
Current Consumption	: 2mA, maximum
Semiconductor Complement	: 2 transistors and 1 diode

1

## 3. PANEL CONTROLS

Frequency dial	: Six frequency scales calibrate according to bands and color of the coil in use.
Meter	: For indicating resonance and for battery check.
Function switch	: Three positions
	OSC : For use as a dip meter or absorption wavemeter.
	MOD : Applies 2kHz AM to the carrier output.
	BATT : For checking condition of CHECK internal battery.
SENSITIVITY	: Adjusts the output level of the oscillator; at OFF, the power supply is cut off and the instrument can be used as an absorption wavemeter.
Earphone jack	: For monitoring AM signals with a crystal earphone (not supplied).

## 4. CIRCUIT DESCRIPTION

The RF oscillator is a Colpitts type using a silicon transistor. The RF voltage is rectified with a diode and the current is indicated on the meter. When the dip meter frequency is in resonance with the external circuit, the meter pointer will dip. The strength of the oscillator output is controlled

with the SENSITIVITY adjuster.

When required, amplitude modulation can be applied to the RF oscillator output. A twin-T type circuit is used for the 2kHz source.

In the absorption wavemeter function, the RF oscillator power is cut off and the tuned and diode-meter circuits are used.

## 5. PRECAUTIONS IN OPERATION

1. During standby periods or when the dip meter is not in use, set SENSITIVITY adjuster at OFF (clicked). This will prolong battery life.
2. Handle the coils carefully when inserting or removing same from the socket.
3. During measurements, the coil should be loosely coupled to the test circuit. The coupling must be as loose as possible to obtain proper meter indication. Close coupling will introduce errors in measurement.

**IMPORTANT!** When checking transmitting equipment, close coupling will damage the diode detector by overload.

4. Disconnect the battery from the internal circuit when the dip meter is not in use for long periods.

3

## 6. OPERATION

### 1. Battery Check and Replacement

1. Set FUNCTION at BATT CHECK.
2. Advance SENSITIVITY to power at on. The meter pointer should swing to the BATT OK portion on the scale. If not, renew the battery.

**NOTE:** To install or replace the battery, remove the covers as follows:

Remove the two screws each at the right and left sides of the instrument; take off the covers. Loosen the battery clamp screw at center of the chassis and take out the battery. Remove the snap terminals from the battery. To install; reverse steps. Tighten the clamp screw and put on the side covers.

3. After this check, set SENSITIVITY at OFF.

### 2. Use as a Dip Meter

1. Plug the coil for the band in use in the socket.
2. Set FUNCTION at OSC.
3. Advance SENSITIVITY past the click point so that the meter pointer swings to about 0.8.

**NOTE:** When the "F" coil is used, there will be some variation in the meter deflection as the frequency dial is rotated. This effect will be present when the coil is not coupled to the test circuit. Care must be taken during use not to confuse this condition with the actual dip caused by the test circuit.

4. Couple the coil loosely to the test circuit and rotate the frequency dial in small steps until a dip is observed on the meter. (When there are two dips close together, place the coil farther from the test circuit.)
5. Adjust the frequency dial until the maximum dip is obtained.
6. Note the frequency on the frequency dial.
7. Set SENSITIVITY at OFF after the tests.
8. Examples of typical coupling methods are shown in Fig. 1.

NOTE: It is important that the power, if applied, to the test circuit is turned off.




Fig. -1 Dip meter coupling methods.

5

A and B : Director or link types.

C : For shielded circuits; the coupling capacitor should be as small as possible.

D : Determination of resonant frequency of an antenna; use a 1-turn coil in the antenna.

(For vertically grounded antennas, do not forget to make the ground connection.)

E : Determination of electrical length of quarter-wave and half-wave coaxial feeders.

### 3. Use as an Absorption Wavemeter

1. Plug the coil for the band in use in the socket.
2. Set SENSITIVITY at OFF.
3. Set FUNCTION at OSC.
4. When the coil is coupled to the coil in an oscillator, or RF tank coil in a transmitter, the meter pointer will swing upward when tuned to resonance. The amount of the swing depends on the strength of the oscillations and/or coupling distance.

NOTE: In checking transmitters, take care that the energy picked up will not overload the instrument, otherwise the diode detector will be damaged. Further, do not touch any of the high voltage points within the transmitter; safety is important.

5. Read the frequency.
6. Typical application are shown in Fig. 2.

For "remote" indication of the modulated output, the meter can be connected to the ear-phone plug; the inner contact is the + side.

#### 6. Use as a Crystal Oscillator

In place of the coil, a quartz crystal, 1 – 15Mhz, in the FT-243 holder, inserted in the socket will produce output at the crystal frequency.

Set FUNCTION at OSC and SENSITIVITY as required for the output. The frequency dial is rotated until the most stable condition is obtained.

Set the dip meter near the receiver for signal pickup.

For modulated output, set FUNCTION at MOD.

#### 7. Audio Signal Output

Set up the instrument for dip meter operation, see Sect. 2, with coil E or F in the socket (However, the RF is not used.)

Set FUNCTION at MOD.

Audio output at approximately 2kHz is available at the PHONE jack. The plug connections are as follows:

Inner contact for the "hot" side and the sleeve for ground.

This signal can be used for checking audio circuits.

#### 8. Inductance and Capacitance Measurements

Unknown inductances and capacitances can be determined with use of the dip meter, see

9

Sect. 5–2 and "A" in Fig. 1.

##### A; Inductance measurement:

A capacitor of known value is required. This capacitor is connected in parallel with the unknown coil and the resonant frequency is noted.

The inductance is calculated from the following:

$$L = \frac{25.330}{C_{pF} (f_{MHz})^2} \mu H$$

With use of a standard 100pF capacitor, the measureable inductance range is from 113μH to 0.004μH with use of the frequency range in the dip meter.

##### B: Capacitance measurement:

An inductance of known value is connected in parallel with the unknown capacitor, and the resonant frequency is noted.

The capacitance is calculated from the following:

$$C = \frac{25,330}{L_{\mu H} (f_{MHz})^2} pF$$